

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100387
Start Date: 11/7/22
Date Received: 11/7/22

Analyst(s): C. O'Bryan

Reason for instrument evaluation

Unable to load ticket database

Visual Inspection

- Clean
- Replace missing screws
- Cermetex Box (modem) is secure
- Ensure modem cable present
- Double Check Valve secured to wall
- All tubing and connections secure
- O-ring inspection and ensure 2 replacement O-rings in bag

As Found Conditions:

Barometric Pressure (inHg): _____ Temperature (C): _____ Relative Humidity (%): _____

Linearity Test (attached)

0.020g/210L Mnf: _____ Lot: _____ Exp: _____
0.080g/210L Mnf: _____ Lot: _____ Exp: _____
0.300g/210L Mnf: _____ Lot: _____ Exp: _____

Notes

- VOC (attached)
- Verify software version
- VOC Internal Standard value $\leq 3.00\%$
- Non-Drinking Subject Test (attached)

Condition of Item: Good Needs Repair

- Ready for Adjustment
- Ready for Calibration

End Date: 11/15/22

*Unable to perform any testing due to pop up bar + instrument restart (unable to load ticket db)
COB 11/15/22*