

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/9/21 Instrument Serial Number: 100365

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Derek Sitz/APD

Describe the Status Change or Issue with the Instrument:

This instrument had an "undefined error" status message on 10/8/21 in the morning. On 10/8/21 in the evening, I had a subject give two different samples where the instrument seemed to take the breath samples as the subject samples came up as a .235 and .207 but then the instrument gave a status message of "Test Failed." When I printed the last result, it came up with a sample from earlier in the evening. When I looked under reports, there was no record for either breath tests. This instruments had a blank error and other errors the day before.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 10/9/21 0030

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740