

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100403

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/04/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.11: 0.078  
LOT #: AG018102  
EXPIRATION: 06/29/2022  
TANK PRESSURE: 229 psi

BLANK TEST	0.000	13:38
INTERNAL STANDARD	VERIFIED	13:38
EXTERNAL STANDARD	0.079	13:39
BLANK TEST	0.000	13:39
EXTERNAL STANDARD	0.079	13:40
BLANK TEST	0.000	13:41
EXTERNAL STANDARD	0.079	13:41
BLANK TEST	0.000	13:42
EXTERNAL STANDARD	0.080	13:42
BLANK TEST	0.000	13:43
EXTERNAL STANDARD	0.080	13:43
BLANK TEST	0.000	13:44

Average = 0.0794  
Std Dev = 0.0005

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C PASSED  
Breath Tube = 42.1°C PASSED

### PUMP INFO

Flow Rate = 4.619 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/3/21

Brandi M. Barnett  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 03 day of 12, 2021



Notary Public  
My Commission Expires With Office



 11/4/21