

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 8/4/2022

Instrument #: 100422

Instrument Location: JBER

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

I contacted JBER Breath Test Supervisor, Scott Gardner, regarding status messages between 2022.05.31 and 2022.07.18. There was an elevated number of RFI status messages (9/37). All were followed by completed tests, but indicate the presence of RFI in the vicinity of the instrument.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: [Click or tap here to enter text.](#)

-----  
**For Use by SCDL**

Additional Notes

[Click or tap here to enter text.](#)

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

**100422 2022.08.04**  
**COB 2022.08.10**