

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100387

Received From: Palmer PD

Date: 11/7/22

Analyst: C. O'Bryant

Reason for instrument return:

Unable to load ticket database

Change instrument and ^{cos 11/7/22} ~~case location~~ in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload Not performed

Update agency information and polling status in DMHost Calling Program

Check for VOCs * Pulled the SD card to ensure all tests on memory were present in DMHost. All files present as of last upload.

From Intoximeters

Intoximeters Repair Paperwork (attached)