

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 2-27-2018 Instrument Serial Number: 100669

Instrument Location: Anchorage Police Department - Anchorage Jail

Supervisor Name and Agency: Ofc. David Noll, Anchorage Police Department

Describe the Status Change or Issue with the Instrument:

Removed from service at the request of Scientific Director Charles Foster due to higher than expected number of Interference Detected status messages.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 2-27-2018/0630 hours

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740