

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100360

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/16/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.24: 0.081  
LOT #: 01315080A3  
EXPIRATION: 02/05/2017  
TANK PRESSURE: 866 psi

BLANK TEST	0.000	15:21
INTERNAL STANDARD	VERIFIED	15:21
EXTERNAL STANDARD	0.080	15:21
BLANK TEST	0.000	15:22
EXTERNAL STANDARD	0.079	15:22
BLANK TEST	0.000	15:23
EXTERNAL STANDARD	0.081	15:23
BLANK TEST	0.000	15:24
EXTERNAL STANDARD	0.080	15:24
BLANK TEST	0.000	15:25
EXTERNAL STANDARD	0.080	15:25
BLANK TEST	0.000	15:26

Average = 0.0800  
Std Dev = 0.0007

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 46.8°C PASSED

### PUMP INFO

Flow Rate = 4.552 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED


INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 9/10/2015  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10<sup>th</sup> day of Sept. 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



  
JMB  
8/31/15