

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100403

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/20/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.87: 0.080
 LOT #: 09523080A1
 EXPIRATION: 05/05/2025
 TANK PRESSURE: 749 psi

BLANK TEST	0.000	13:44
INTERNAL STANDARD	VERIFIED	13:44
EXTERNAL STANDARD	0.081	13:44
BLANK TEST	0.000	13:45
EXTERNAL STANDARD	0.081	13:45
BLANK TEST	0.000	13:46
EXTERNAL STANDARD	0.082	13:46
BLANK TEST	0.000	13:47
EXTERNAL STANDARD	0.081	13:47
BLANK TEST	0.000	13:48
EXTERNAL STANDARD	0.082	13:48
BLANK TEST	0.000	13:49

Average = 0.0814
 Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
 PIC: 3.02
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 45.8°C	PASSED

PUMP INFO

Flow Rate = 4.657 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED


FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------


I, Derek J. Walton, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) As of the date signed below, I certify that the calibration of the referenced instrument was accurate on the date in which the above tests were performed and therefore certified for evidentiary use in the State of Alaska.



 Derek J. Walton
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 10 day of 08, 2023



 Notary Public
 My Commission Expires With Office



OK 7/31/23