

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100682

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/16/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 610 psi

BLANK TEST	0.000	11:58
INTERNAL STANDARD	VERIFIED	11:58
EXTERNAL STANDARD	0.077	11:58
BLANK TEST	0.000	11:59
EXTERNAL STANDARD	0.077	11:59
BLANK TEST	0.000	12:00
EXTERNAL STANDARD	0.077	12:00
BLANK TEST	0.000	12:01
EXTERNAL STANDARD	0.077	12:02
BLANK TEST	0.000	12:02
EXTERNAL STANDARD	0.078	12:03
BLANK TEST	0.000	12:04

Average = 0.0772
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 45.3°C	PASSED

PUMP INFO

Flow Rate = 4.654 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO


Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 7/25/19

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of 07, 20 19



Notary Public
My Commission Expires With Office



COB 7/15/19