

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100691

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/05/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.02: 0.080
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 944 psi

BLANK TEST	0.000	10:21
INTERNAL STANDARD	VERIFIED	10:22
EXTERNAL STANDARD	0.081	10:22
BLANK TEST	0.000	10:23
EXTERNAL STANDARD	0.081	10:23
BLANK TEST	0.000	10:24
EXTERNAL STANDARD	0.081	10:24
BLANK TEST	0.000	10:25
EXTERNAL STANDARD	0.081	10:25
BLANK TEST	0.000	10:26
EXTERNAL STANDARD	0.081	10:26
BLANK TEST	0.000	10:27

Average = 0.0810
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 44.7°C PASSED

PUMP INFO

Flow Rate = 4.742 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


4/9/21

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of 4, 2021


Notary Public
My Commission Expires With Office



DM 02/05/21