

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/15/2022

Instrument #: 100671

Instrument Location: Metlakatla, Alaska

Supervisor Name and Agency: Bruce Janes, Metlakatla Police Department

Describe the Status Change or Issue with the Instrument:

The touch screen is not working

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 11/15/2022 at 1200 hours

-----  
**For Use by SCDL**

**Additional Notes**

Click or tap here to enter text.  
Troubleshooting included restarting the instrument, which did not fix the screen issue. CKG 11/16/2022

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

Page 1 of 2

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

For questions contact the Breath Alcohol Program at 907-269-5740

Page 2 of 2

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor