

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 11-23-21 Instrument #: 100376

Instrument Location: WASILLA PD

Supervisor Name and Agency: JASON C. HOLMGREN WPD

Describe the Status Change or Issue with the Instrument:

STATE CRIME LAB ISSUED NEW INSTRUMENT

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 11-23-21 / 0225

For Use by SCDL

Additional Notes

Removed from service due to internal
Standard errors. CD3 11/24/21

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740