

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100384

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/25/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.54: 0.079  
LOT #: AG634001  
EXPIRATION: 12/05/2018  
TANK PRESSURE: 584 psi

BLANK TEST	0.000	20:51
INTERNAL STANDARD	VERIFIED	20:51
EXTERNAL STANDARD	0.072	20:51

Standard Out of Range

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 44.6°C	PASSED

### PUMP INFO

Flow Rate = 4.714 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED
--------