

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 08/08/2017
Effective: 08/10/2017

Version: DMEF2017 R1
Status: Active

Instrument # 100345

Date: 12/6/17

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Intermittent status messages (blank error, interference, etc)

2. Visual Inspection (internal and external)

3. O-ring inspection

4. Troubleshooting/Repair

No status messages during evaluation
VOC, will monitor during certification.

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100345

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/06/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.71: 0.079
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 176 psi

BLANK TEST	0.000	15:28
INTERNAL STANDARD	VERIFIED	15:28
Xq = 0.093	(0.28%)	
EXTERNAL STANDARD	0.076	15:29
X[1] = 0.0761	(-0.0033) (0.0018)	
BLANK TEST	0.000	15:30
EXTERNAL STANDARD	0.076	15:30
X[1] = 0.0764	(-0.0033) (0.0012)	
BLANK TEST	0.000	15:31
EXTERNAL STANDARD	0.076	15:31
X[1] = 0.0764	(-0.0037) (0.0014)	
BLANK TEST	0.000	15:32
EXTERNAL STANDARD	0.077	15:32
X[1] = 0.0767	(-0.0032) (0.0019)	
BLANK TEST	0.000	15:33
EXTERNAL STANDARD	0.077	15:33
X[1] = 0.0766	(-0.0037) (0.0014)	
BLANK TEST	0.000	15:34

Average = 0.0764
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 48.1°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.67 V	PASSED
(0.97 V - 2.25 V)	
Cooler Voltage = 1.74 V	PASSED
(1.03 V - 2.39 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 525 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.7 in	

PUMP INFO

Flow Rate = 4.251 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD PASSED

Xq = 0.092 (0.66% - 4.00%)