

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/01/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 46.5°C PASSED

PUMP INFO

Flow Rate = 4.428 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

CB 8/10/21

WLB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/01/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 48.1°C

PASSED

PUMP INFO

Flow Rate = 4.497 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

CS 8/10/21

AMB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/07/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 41.1°C

PASSED

PUMP INFO

Flow Rate = 4.317 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

COB 8/10/21
OUB
7/30/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/12/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.185 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

Diagnostic Check Failed FAILED

003 8/10/21
0008
7/31/2021

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/16/2021

External Standard Test Values

Diagnostic Check

VERSIONS	
DMT: 3.02	
PIC: 3.02	
Modem: 2.6	
Questions: 2.2	
TEMPERATURES	
Sample Chamber = 49.0°C	PASSED
Breath Tube = 46.2°C	PASSED
PUMP INFO	
Flow Rate = 4.437 L/M	PASSED
DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED
FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED
INTERNAL STANDARD	FAILED
Diagnostic Check Failed	

08/10/21
AUS
7/30/21

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/16/2021

External Standard Test Values

Diagnostic Check

VERSIONS	
DMT: 3.02	
PIC: 3.02	
Modem: 2.6	
Questions: 2.2	
TEMPERATURES	
Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.1°C	PASSED
PUMP INFO	
Flow Rate = 4.395 L/M	PASSED
DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED
FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED
INTERNAL STANDARD	FAILED
Diagnostic Check Failed	

CCB 8/10/21

CCB
7/30/2021