

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Issued: 4/12/2018
Effective: 4/13/2018

Version: DMCAF 2018 R0
Status: Active

Instrument # 100347

Analyst Derek Walton

Date: 06/21/2019

Reason for Calibration

Internal standard >3.00% on evaluation

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Qoth
- c. Lot 18370
- d. Expiration Date 12/05/2020

2. Wet Bath Simulator

- a. Serial Number SD1198
- b. Date Calibrated 03/06/2019

3. Attach Calibration Factors

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100347

Date: 06/21/2019
Time: 10:21:07

OPERATOR NAME:
DEREK WALTON
CERTIFICATE #: 9508

Ca = 0.1000
CAL = 0.967974 0.800 <= CAL < 1.200
b1 = 0.0003 0.0000 <= b1 < 0.0040
b2 = 0.0042 0.0010 <= b2 < 0.0100
b3 = 0.0004 0.0000 <= b3 < 0.0040
Xq = 0.0905 0.0500 <= Xq < 0.2000
a21 = 1.182505 1.050 <= a21 < 1.300
a31 = 0.445320 0.300 <= a31 < 0.800