

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 1/4/2013
Effective: 1/4/2013

Version: CIISF2013 R0
Status: Active

Date: 5 Nov 18 Instrument Serial Number: 100689

Instrument Location: Kotzebue Police Dept.

Supervisor Name and Agency: Sgt. DAVID ROBERTS, (KA) KOTZEBUE POLICE DEPT.

Describe the Status Change or Issue with the Instrument:

Spoke @ crime lab who advised to take it out of service & new one being sent. Failed VOC's, Interferences, blank runs, & tube temps.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: 31 Oct 18 @ 08:30 hrs

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740