

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 10/6/2022

Instrument #: 100684

Instrument Location: Sitka, AK

Supervisor Name and Agency: Nathan Duce, Alaska State Troopers

Describe the Status Change or Issue with the Instrument:

Datamaster DMT wouldn't run and when it was rebooted it displayed the following status messages; Undefined Error, Test Failed, and DMT restarted due to fatal error.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 10/6/2022 no longer using for training

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor