

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100370

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/14/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.00: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1193 psi

BLANK TEST	0.000	08:17
INTERNAL STANDARD	VERIFIED	08:17
EXTERNAL STANDARD	0.080	08:18
BLANK TEST	0.000	08:19
EXTERNAL STANDARD	0.080	08:19
BLANK TEST	0.000	08:20
EXTERNAL STANDARD	0.081	08:20
BLANK TEST	0.000	08:21
EXTERNAL STANDARD	0.080	08:21
BLANK TEST	0.000	08:22
EXTERNAL STANDARD	0.080	08:22
BLANK TEST	0.000	08:23

Average = 0.0802
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 45.6°C PASSED

PUMP INFO

Flow Rate = 4.771 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 6/25/18

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of June, 2018


Carlie K. Bailey, Notary Public
My Commission Expires With Office




OMB
6/6/18