

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100668

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/16/2022

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.70: 0.082
LOT #: AG135004
EXPIRATION: 12/16/2023
TANK PRESSURE: 768 psi

BLANK TEST	0.000	08:18
INTERNAL STANDARD	VERIFIED	08:19
EXTERNAL STANDARD	0.083	08:19
BLANK TEST	0.000	08:20
EXTERNAL STANDARD	0.083	08:20
BLANK TEST	0.000	08:21
EXTERNAL STANDARD	0.082	08:21
BLANK TEST	0.000	08:22
EXTERNAL STANDARD	0.083	08:22
BLANK TEST	0.000	08:23
EXTERNAL STANDARD	INTERFERENCE	08:23

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 46.6°C	PASSED

PUMP INFO

Flow Rate = 4.995 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only dmt 01/05/23