

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100361

Received From: APD

Date: 11/19/2021

Analyst: Curlye Flaister

Reason for instrument return:

Breath hose not heating to temperature.

Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database

no case, local agency
Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)