

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100690

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:05/30/2019

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 48.0°C PASSED

### PUMP INFO

Flow Rate = 4.187 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED

PUMP OFF FAILED

### FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

### INTERNAL STANDARD

PASSED

Diagnostic Check Failed