

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100412

Received From Unalaska DPS

Date 02/17/21

Analyst Derek Walter

1. Reason for instrument return:

Internal Standard errors

2. Change instrument and case location in JusticeTrax Breath Alcohol Database
3. Ensure change in instrument status form is in JusticeTrax
4. Upload
5. Update agency information and polling status in DM Host Calling Program
6. Check for VOCs

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)