

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100255

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:07/23/2018

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 2.05

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 47.4°C

PASSED

### PUMP INFO

Flow Rate = 4.576 L/M

PASSED

User Abort