

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 3/14/2023

Instrument #: 100689

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Derek Sitz, Anchorage Police Dept.

Describe the Status Change or Issue with the Instrument:

Keyboard would not work. Tried resetting the Datamaster, using different USB ports, verified keyboard worked in other Datamaster, and tried a different keyboard. Replaced Datamaster with 100667 and took 100689 out of service. Ran VOC on 100667.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 3/14/23 2330

-----  
**For Use by SCDL**

Additional Notes

[Click or tap here to enter text.](#)

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor