

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100699

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/18/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.02: 0.080  
LOT #: AG018102  
EXPIRATION: 06/29/2022  
TANK PRESSURE: 942 psi

BLANK TEST                      0.000    12:02  
INTERNAL STANDARD            ERROR    12:02

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C      PASSED  
Breath Tube      = 46.6°C      PASSED

### PUMP INFO

Flow Rate = 4.586 L/M      PASSED

### DETECTOR INFO

PUMP ON                      PASSED  
PUMP OFF                      PASSED

### FILTER INFO

Filter 1                      PASSED  
Filter 2                      PASSED  
Filter 3                      PASSED

INTERNAL STANDARD      PASSED

CAB 8/10/21  
CAB  
7/21/2021