

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.436 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

CSB 8/10/21

CSB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

Diagnostic Check

VERSIONS	
DMT: 3.02	
PIC: 3.03	
Modem: 2.6	
Questions: 2.2	
TEMPERATURES	
Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED
PUMP INFO	
Flow Rate = 4.461 L/M	PASSED
DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED
FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED
INTERNAL STANDARD	FAILED
Diagnostic Check Failed	

CAB 8/10/21
CUB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.60: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 911 psi

BLANK TEST	0.000	11:12
INTERNAL STANDARD	VERIFIED	11:12
EXTERNAL STANDARD	0.080	11:12
BLANK TEST	0.000	11:13
EXTERNAL STANDARD	INTERFERENCE	11:13

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.493 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

008/10/21

008
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.60: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 903 psi

BLANK TEST	0.000	11:18
INTERNAL STANDARD	VERIFIED	11:18
EXTERNAL STANDARD	0.076	11:18
BLANK TEST	0.000	11:19
EXTERNAL STANDARD	0.079	11:20
BLANK TEST	0.000	11:21
EXTERNAL STANDARD	0.080	11:21
BLANK TEST	0.000	11:22
EXTERNAL STANDARD	0.079	11:22
BLANK TEST	0.000	11:23
EXTERNAL STANDARD	INTERFERENCE	11:23

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.574 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

003 8/10/21

003
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

Diagnostic Check

VERSIONS	
DMT: 3.02	
PIC: 3.03	
Modem: 2.6	
Questions: 2.2	
TEMPERATURES	
Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED
PUMP INFO	
Flow Rate = 4.492 L/M	PASSED
DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED
FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED
INTERNAL STANDARD	FAILED
Diagnostic Check Failed	

COB 8/10/21

CLB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/06/2021

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO
Flow Rate = 4.570 L/M PASSED

DETECTOR INFO
PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO
Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD FAILED

Diagnostic Check Failed

003 8/10/21

003
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Serial #: 100399

Date: 07/06/2021

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO
Flow Rate = 4.584 L/M PASSED

DETECTOR INFO
PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO
Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD FAILED

Diagnostic Check Failed

COB 8/10/21

WUB
7/21/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/07/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 48.1°C

PASSED

PUMP INFO

Flow Rate = 4.517 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

COB 8/10/21
OUB
7/30/2021

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/16/2021

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C

PASSED

Breath Tube = 48.1°C

PASSED

PUMP INFO

Flow Rate = 4.630 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

COB 8/10/21

COB
7/30/2021