

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 12/28/18 Instrument Serial Number: 100678 / 100667 *brub 12/28/18*

Instrument Location: AST/Mat-Su West

Supervisor Name and Agency: Trooper Lee Phillips AST

Describe the Status Change or Issue with the Instrument:

As mentioned in previous email, multiple "Detector overflow" status messages on previous instrument S/N 100667.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Instrument S/N 100678 placed IN SERVICE on 12/28/18 @ 0500 hours.

Out of Service Date/Time: Instrument S/N 100667 OUT OF SERVICE on 12/28/18 @ 0400 hours.

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740