

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100691

Received From Valdez PD

Date 12/10/2019

Analyst Derek Walton

1. Reason for instrument return:

Preventive maintenance, standard out of range, internal standard error

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in ~~JusticeTrax~~ *on network drive DM 12/10/19*

4. Upload NA - ^①Instrument won't turn on - see evaluation *DM 12/10/19*

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs NA - Sec ①

From Intoximeters

7. Intoximeters Repair Paperwork (attached)