

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100694

Received From APD

Date 09/08/20

Analyst Derek Walton

1. Reason for instrument return:

Blank error, internal standard error, standard out of range

2. Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database

Not returned with case DM 09/08/20

3. Ensure change in instrument status form is in JusticeTrax

4. Upload ^① Unit won't turn on DM 09/08/20 - see evaluation form 9/8/20

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs ^①

From Intoximeters

7. Intoximeters Repair Paperwork (attached)