

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100699

Received From: Valdez PD

Date: 5/16/24

Analyst: Megan Crouse

*Supervised casework
with Derek Walton 05/16/24
JM*

Reason for instrument return:

Several standard out of range messages. The external
standard is also reading low.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)