

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 3-7-15 Instrument Serial Number: 100381

Instrument Location: Anchorage Jail north room

Supervisor Name and Agency: Derek Site APO

Describe the Status Change or Issue with the Instrument:

The instrument was only using diagnostic and internal checks with no external standard checks.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 3-7-15 0500

For Use by SCDL

Additional Notes

This instrument was used for training at the laboratory. The options for the external standard before and after the subject sample were turned off for training purposes. This option was inadvertently left off after BMB when the instrument was sent to APD. The options were reset and the instrument was placed back in service on 3/10/15. A non-drinking subject test was performed to demonstrate it was working properly. BMB 3/12/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740