

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/20/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.99: 0.080
LOT #: 09214080A2
EXPIRATION: 05/01/2016
TANK PRESSURE: 971 psi

BLANK TEST	0.000	03:05
INTERNAL STANDARD	VERIFIED	03:05
EXTERNAL STANDARD	0.080	03:06
BLANK TEST	0.000	03:07
EXTERNAL STANDARD	0.081	03:07
BLANK TEST	0.000	03:08
EXTERNAL STANDARD	0.081	03:08
BLANK TEST	0.000	03:09
EXTERNAL STANDARD	0.081	03:09
BLANK TEST	0.000	03:10
EXTERNAL STANDARD	0.081	03:10
BLANK TEST	0.000	03:11

Average = 0.0808
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.01
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 39.0°C PASSED

PUMP INFO

Flow Rate = 4.386 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

9/10/2015

Subscribed and sworn before me this 10th day of Sept., 20 15


Nikki Roth, Notary Public
My Commission Expires With Office



BMB
9/10/15