

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100349

Received From Intoximeters

Date 09/30/2019

Analyst Derek Walton

1. Reason for instrument return:

Factory refurbish - screen not functioning correctly,  
pump failure, fatal system errors.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax  NA

4. Upload  NA

5. Update agency information and polling status in DM Host Calling Program  NA

6. Check for VOCs  NA

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)

**SRO Number:** SRO-036628

**Received Date:** 09/16/2019

**Completion Date:** 09/24/2019

**SRO Type:** REPAIR

**SRO Description:** DMT Repair/Refurb

**Customer ID:** C000003651

**Customer PO:** 09162019-Barnett

**Contact:**

**Email:**

**Ship To:**

**Units on SRO**

Alaska Scientific Crime Detection Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

1 LEGDMT LEGACY DMT

<b>Service Line: 1</b>		<b>**All Instruments are Calibrated to Factory Specifications**</b>
<b>Unit Type:</b>	(LEGDMT) LEGACY DMT	
<b>Serial Number:</b>	50100349	
<b>Reason:</b>	DMT Functional	
	No Displ	
<b>Reason Notes:</b>	Customer Reported Issue: Display not working	
	Perform refurbishment and regulator replacment	
<b>Resolution:</b>	Replaced	
	RPL PCB	
<b>Resolution Notes:</b>	90-8000-00 Prev Maint Contract p/n Replaced dry gas regulator Replaced breath tube clip Replaced SI board Installed Alaska software Rebuilt filter assembly and added epoxy to motor wires Optimized detector Adjusted tech screen settings and RFI sensitivity Added o-ring to end of breath tube  Parts used: 001217-00 regulator 59-0036-00 Breath tube clip 59-0022-00 Washer countersink #6 59-0000-08 Cable tie 001199-00 3.3voltage regulator  Calibration solution (0.200) Lot#19040 Bottle#78 Dry gas: AG916303, Tank 014, Exp. 12 Jun 2021	

<b>Parts:</b>	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	59-0036-00	DMT HRDW CLIP BREATH TUBE HOLD 37058
	1	001217-00	DMT HRDW GAS MOUNT ASSY W/VALVE
	1	59-0022-00	DMT HRDW WASHER CNTRSINK #6 SS 34330
	1	59-0000-08	DMT HRDW CABLE TIE HOLDER 30320
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN
	1	90-8000-00	CONTRACT PREV MAINT DMT

# Order Packing Slip Report

**Intoximeters**  
Experience • Service • Integrity

9/25/2019 2:22:41 PM

**Packing Slip:** 186864

**Pack Date:** 9/25/2019

**SRO:** SRO-036628



**From:** Intoximeters, Inc  
2081 Craig Road  
St Louis MO 63146  
USA

**Customer:** C000003651

**Ship To:** 1

Alaska Department of Public Safety  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Alaska Scientific Crime Detection  
Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

**Customer PO:** 09162019-Barnett

**Ship Via:** FDX

FED EX

**Weight:** 38.00

**Delivery Terms :** 2D

**FOB :** ORIGIN

**# Packages:** 1

**Customer Ship Account :**

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100349
1	59-0036-00	EA		1.00	0.00
	DMT HRDW CLIP BREATH TUBE HOLD 37058				
2	001217-00	EA		1.00	0.00
	DMT HRDW GAS MOUNT ASSY W/VALVE				
3	59-0022-00	EA		1.00	0.00
	DMT HRDW WASHER CNTRSNK #6 SS 34330				
4	59-0000-08	EA		1.00	0.00
	DMT HRDW CABLE TIE HOLDER 30320				
5	001199-00	EA		1.00	0.00
	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN				
6	90-8000-00	EA		1.00	0.00
	CONTRACT PREV MAINT DMT				