

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 8-26-15 Instrument Serial Number: 100378

Instrument Location: Anchorage Jail - South Room

Supervisor Name and Agency: David Noll - APD

Describe the Status Change or Issue with the Instrument:

Shows no flow rate with exhalation

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 8-26-15 / 0430

For Use by SCDL

Additional Notes

Officer Noll provided a sample and his breath flow was not registering. (BB 8/26/15)

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100378

Date: 08/26/2015
Time: 04:28:23

SUBJECT NAME:
TEST TEST
LICENSE #: 1234567

OPERATOR NAME:
DAVID M NOLL
OPERATOR'S #: 6520

DEPT/AGENCY:
ANC1
CASE/REPORT: 15-0
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.10: 0.080
LOT #: 09214080A2
EXPIRATION: 05/01/2016

DIAGNOSTIC CHECK PASSED 04:29
BLANK TEST 0.000 04:29
INTERNAL STANDARD VERIFIED 04:29
EXTERNAL STANDARD 0.081 04:29
BLANK TEST 0.000 04:30
SUBJECT SAMPLE INCOMPLETE 04:32

NON
Drinking
Test

Alcohol (g/100L)
Flow Rate(L/M)
Min. Flow Rate (L/M)

