

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 6/13/19 Instrument Serial Number: 100355

Instrument Location: Seward Community Jail

Supervisor Name and Agency: Sgt. Schaefermeyer Seward Police Department

Describe the Status Change or Issue with the Instrument:

Swapped for new unit

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 6/13/19 12:15 pm

For Use by SCDL

Additional Notes

Internal standard errors, standard out of range.
JM 06/24/2019

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740