

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 3/17/15 Instrument Serial Number: 100423

Instrument Location: Ft. Wainwright

Supervisor Name and Agency: Brandi Barnett / SCDL

Describe the Status Change or Issue with the Instrument:

I was contacted by Sgt Atkins on Saturday morning after they received two fatal system errors in a row. He was instructed to try and reboot again and if that did not work to contact me and we would trouble shoot further. I contacted him Monday and he stated that after the second reboot the instrument was working again. He was instructed to let us know if he had any further problems.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740