

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100697

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/14/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.20: 0.078  
LOT #: 09214080A3  
EXPIRATION: 05/01/2016  
TANK PRESSURE: 317 psi

BLANK TEST	0.000	19:57
INTERNAL STANDARD	VERIFIED	19:57
EXTERNAL STANDARD	0.078	19:57
BLANK TEST	0.000	19:58
EXTERNAL STANDARD	0.078	19:58
BLANK TEST	0.000	19:59
EXTERNAL STANDARD	0.077	19:59
BLANK TEST	0.000	20:00
EXTERNAL STANDARD	0.078	20:00
BLANK TEST	0.000	20:01
EXTERNAL STANDARD	0.078	20:01
BLANK TEST	0.000	20:02

Average = 0.0778  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 4.118 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

10/7/2015

Subscribed and sworn before me this 7<sup>th</sup> day of Oct, 2015

  
Nikki Roth, Notary Public  
My Commission Expires With Office



008 9/24/15