

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100386

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/31/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.38: 0.079  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 1197 psi

BLANK TEST	0.000	17:01
INTERNAL STANDARD	VERIFIED	17:01
EXTERNAL STANDARD	0.076	17:01
BLANK TEST	0.000	17:02
EXTERNAL STANDARD	INTERFERENCE	17:03

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 42.1°C	PASSED

### PUMP INFO

Flow Rate = 4.675 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED