

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100364

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/26/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.19: 0.081  
LOT #: 09214080A2  
EXPIRATION: 05/01/2016  
TANK PRESSURE: 809 psi

BLANK TEST	0.000	09:53
INTERNAL STANDARD	VERIFIED	09:53
EXTERNAL STANDARD	0.076	09:53
BLANK TEST	0.000	09:54
EXTERNAL STANDARD	0.077	09:54
BLANK TEST	0.000	09:55
EXTERNAL STANDARD	0.077	09:55
BLANK TEST	0.000	09:56
EXTERNAL STANDARD	0.077	09:57
BLANK TEST	0.000	09:57
EXTERNAL STANDARD	0.077	09:58
BLANK TEST	0.000	09:59

Average = 0.0768  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.6°C PASSED  
Breath Tube = 48.0°C PASSED

### PUMP INFO

Flow Rate = 5.078 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD


PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1<sup>st</sup> day of June, 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



COB 5/27/15