

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 8-1-15 Instrument Serial Number: 100397

Instrument Location: Anchorage Jail

Supervisor Name and Agency: JEREMY EURIE - APD

Describe the Status Change or Issue with the Instrument:

DATAMASTER Displaying "Incomplete" status message  
Not registering breath/volume/flow

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 8-1-15 0332 hrs.

For Use by SCDL

Additional Notes

5-way valve appears to be sticking. COB 8/3/15

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740