

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100413

Received From Kenai PD

Date 03/18/2021

Analyst Derek Walton with
Carlise Bailey - trainee

1. Reason for instrument return:

Tank interference, simulator timeout status messages

- 2. Change instrument and case location in JusticeTrax Breath Alcohol Database
- 3. Ensure change in instrument status form is in JusticeTrax
- 4. Upload
- 5. Update agency information and polling status in DM Host Calling Program
- 6. Check for VOCs

From Intoximeters

- 7. Intoximeters Repair Paperwork (attached)