

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100666

Received From NPAS

Date 10/27/15

Analyst CTF

1. Reason for instrument return:

NPAS Repair due to Message pop-ups

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax N/A

4. Upload N/A

5. Update agency information and polling status in DM Host Calling Program N/A

6. Check for VOCs N/A

From NPAS

7. NPAS Repair Paperwork (attached)

Packing List

From: National Patent Analytical Systems Inc.
2090 Harrington Mem. Rd., Mansfield, Ohio 44903

Cust No. AK6056A
Sales Order No. 49974

Ord Date 9/23/15

Ship Date: 10/13/15
Packing List No. 74598

Cust Purch Ord No. COLLEEN

Ship To:

Alaska Dept of Public Safety
Scientific Crime Detection
4805 Dr. Martin Luther King Dr.

Anchorage Ak 99507

Attn:

Comments Service for 100666

Bill To:

Alaska Dept of Public Safety
Scientific Crime Detection
4805 Dr Martin Luther King Dr.

Anchorage Ak 99507 Freight

25.00

Ship Via FedEx

Part / Drawing No / Rev No	Quantity Ordered	This Shipment	Back Ordered	Total Ship To Date	Description	NP Invent Ref
No Draw # / NONE 100666 Serial Nos.	1.0	1		1.0	DATAMASTER DMT SERVICE RETURN	9DMT
No Draw # / NONE Serial Nos.	1.0	1		1.0	NANUK CASE	NOTE
No Draw # / NONE warranty Serial Nos.	2.0	2		2.0	Service Charge for Repair	r500-1

Work Ord No 49974

National Patent Analytical Systems, Inc.
Service Action Report

Serial No 100666

Date 9/23/15

Customer Alaska Dept of Public

Phone 907-269-5582

Contact Bolz

War Exp 9/26/13

Mx Exp

Ship Damage? Y N

Cust Pick Up?? Y N

Initial Comments: note any shipping damage or estimates required, along with accessories

Describe problem seen and/or attach documents received

see attached document

Keyword

Finding

Error Code

Enter Hours

Repair Comments and work Performed

updated software, recalibrated and ran several accuracy tests with no problems

Solution Lot No.

Voltage Set Values In		Voltage Set Values Out	
Lamp	Brth Vol	Lamp	Brth Vol
Meter	Filter Dif	Meter	Filter Dif
Stability		Stability	

Part No	Qty Ord	Pick	Back Ordered	Total shin	Description	Price Ea
9DMT	1.0				DATAMASTER DMT SERVICE RETURN 100666	
NOTE	1.0				NANUK CASE	
r500-1	2.0				Service Charge for Repair warranty	0.000

Date Completed 10/12/15

Technician M. BARBELY

Pick up Signature