

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 9/23/2022

Instrument #: 100687

Instrument Location: Galena Police Department

Supervisor Name and Agency: Robert Nunley Alaska State Troopers

Describe the Status Change or Issue with the Instrument:

Taken out of service due to continued status messages:  
Interference Detected  
Internal Standard Error  
Blank Error

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 8/24/21 1225 hours

Out of Service Date/Time: 9/23/22 1120 hours

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor