

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100360

Received From: Ft. Wainwright

Date: 12/20/23

Analyst: C. O. Bryant

Reason for instrument return:

Instrument out of service since 7/11/23
for multiple status messages - blank errors,
static analysis, filter wheel, etc.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)