

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 3.0

Effective: 6/27/2022

Date: 7/3/2022

Instrument #: 100387

Instrument Location: Palmer PD

Supervisor Name and Agency: Sgt. Szipszky, Palmer Police

Describe the Status Change or Issue with the Instrument:

Initial set up and placed into service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 7/3/22 11:00 AM

Out of Service Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Sent as a replacement Datamaster. Old DMT is experiencing RFI issues. TAL 07/05/2022

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740