

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100680

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/30/2023

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C                      PASSED  
Breath Tube        = 44.2°C                      PASSED

PUMP INFO  
Flow Rate = 4.530 L/M                          PASSED

DETECTOR INFO  
PUMP ON    FAILED  
PUMP OFF     PASSED

FILTER INFO  
Filter 1    PASSED  
Filter 2    PASSED  
Filter 3    PASSED

INTERNAL STANDARD                              PASSED

Diagnostic Check Failed

File Only DW 11/22/23