

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 3.0

Effective: 6/27/2022

Date: 7/4/2022

Instrument #: 100351

Instrument Location: Fairbanks Police Department

Supervisor Name and Agency: Officer Lane Bonham Fairbanks PD

Describe the Status Change or Issue with the Instrument:

Operator notified DataMaster Supervisor that the DataMaster was not taking air in and sounds like it is getting blown into the compartment. I tested the DataMaster and could hear the air being blown into the compartment. I tested the DataMaster again and he accepted a sample. DataMaster has still been taking out of service.

More than likely the 5 way valve got stuck open.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 7/4/22 0800

For Use by SCDL

Additional Notes

A replacement DataMaster has been sent. TAL 07/05/2022

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740