

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 4/21/2016 Instrument Serial Number: 100677  
Instrument Location: Betel Police Department  
Supervisor Name and Agency: Trp. Nich Hayes / AST

Describe the Status Change or Issue with the Instrument:

Returned to errors

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 4/21/2016 / 2105

For Use by SCDL	Additional Notes
	<u>Removed from service for various messages including "unable to load ticket" and "undefined error". See email attached for documentation. CSB 6/15/16</u>

Email completed form to \_\_\_\_\_

For questions contact the Breath Alcohol Program at 907-269-5740

## **Bethel PD Instrument - OS**

O'Bryant, Colleen S (DPS)

**Sent:** Wednesday, April 13, 2016 8:30 AM

**To:** Hayes, Nicolas C (DPS)

**Cc:** adavis@cityofbethel.net; Foster, Charles R (DPS); Roth-Frey, Nicolette A (DPS)

**Nick,**

**Sgt Davis called and said the Bethel PD is down. She has received numerous “unable to load ticket” messages as well as “undefined error”. She is also unable to get the instrument to register flow while providing a non-drinking subject test.**

**We will be sending BPD a new instrument. There isn't currently a breath test supervisor for BPD, are you willing to take their old instrument out of service and install the new one along with the paperwork associated with that?**

**Amy has indicated that she will be requesting permission to attend the breath test supervisor school this fall (Aug 30-Sept 1).**

**Colleen O'Bryant – Forensic Scientist III  
Blood and Breath Alcohol  
Alaska Scientific Crime Detection Laboratory  
4805 Dr. Martin Luther King Jr. Ave  
Anchorage, AK 99507**

**907-269-5783 (direct)**

**907-350-2111 (cell)**