

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/18/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.95: 0.080

LOT #: AG436502

EXPIRATION: 12/30/2026

TANK PRESSURE: 1186 psi

BLANK TEST 0.000 12:02

INTERNAL STANDARD VERIFIED 12:02

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED

Breath Tube = 43.7°C PASSED

PUMP INFO

Flow Rate = 4.710 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only
CW 10/9/25